

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0012
Location: LAKE HOPATCONG SP
Calibration File No.: 00007 Calib. Date: 06/16/2008 Calib. No.: 00002
Certification File No.: 00008 Cert. Date: 06/16/2008 Cert. No.: 00001
Linearity File No.: 00009 Lin. Date: 06/16/2008 Lin. No.: 00001
Solution File No.: 00011 Soln. Date: 07/28/2008 Soln. No.: 00003
Sequential File No.: 00011 File Date: 07/28/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXD S3-0192
Control Solution %: 0.100% Expires: 01/09/2010
Solution Control Lot: 08A049 Bottle No.: 1134

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:43D		
Control 1 EC	0.102%	11:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		
Control 2 EC	0.101%	11:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		
Control 3 EC	0.100%	11:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHNEZLER

First Name: EDWARD

MI: J

Signature: TPRE E. J. Schnezler #5515

Badge No.: 5515

Date: 07/28/2008

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0012
Location: LAKE HOPATCONG SP
Calibration File No.: 00007 Calib. Date: 06/16/2008 Calib. No.: 00002
Certification File No.: 00008 Cert. Date: 06/16/2008 Cert. No.: 00001
Linearity File No.: 00009 Lin. Date: 06/16/2008 Lin. No.: 00001
Solution File No.: 00012 Soln. Date: 08/21/2008 Soln. No.: 00004
Sequential File No.: 00012 File Date: 08/21/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXD S3-0192
Control Solution %: 0.100% Expires: 01/09/2010
Solution Control Lot: 08A049 Bottle No.: 1130

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:07D		
Control 1 EC	0.103%	19:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	19:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:09D		
Control 2 EC	0.102%	19:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:10D		
Control 3 EC	0.101%	19:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHNEZLER

First Name: EDWARD

MI: J

Signature:

PR. I.E.J. Schnezler # 5515

Badge No.: 5515

Date: 08/21/2008

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0012
Location:	NJSP LAKE HOPATCONG	Calib. No.:	00003
Calibration File No.:	00013	Cert. Date:	12/11/2008
Certification File No.:	00014	Lin. Date:	12/11/2008
Linearity File No.:	00015	Soln. Date:	12/11/2008
Solution File No.:	00016	File Date:	12/11/2008
Sequential File No.:	00016		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXD S3-0192
Solution Control Lot:	08A049	Expires:	01/09/2010
		Bottle No.:	1128

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:07S		
Control 1 EC	0.102%	16:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08S		
Control 2 EC	0.102%	16:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:10S		
Control 3 EC	0.102%	16:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	16:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDXDS3-0192 (CAR)

Changed By:

Last Name: ROUSHINKO

First Name: CRAIG

MI: A.

Signature:

TR II C-R #6046

Badge No.: 6046

Date: 12/11/2008

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0012
Location:	NJSP LAKE HOPATCONG	
Calibration File No.:	00013	Calib. Date: 12/11/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 12/11/2008
Solution File No.:	00017	Cert. No.: 00002
Sequential File No.:	00017	Lin. Date: 12/11/2008
		Lin. No.: 00002
		Soln. Date: 02/01/2009
		Soln. No.: 00006
		File Date: 02/01/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXD S3-0192
Solution Control Lot:	08A049	Expires: 01/09/2010
		Bottle No.: 1135

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:38S		
Control 1 EC	0.105%	07:38S	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.104%	07:38S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	07:39S		
Control 2 EC	0.105%	07:40S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.103%	07:40S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:41S		
Control 3 EC	0.104%	07:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHNEZLER

First Name: EDWARD

MI: J

Signature: *TRIE J. Schnezler* #3375

Badge No.: 5515

Date: 02/01/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0012
Location:	NJSP LAKE HOPATCONG	
Calibration File No.:	00013	Calib. Date: 12/11/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 12/11/2008
Solution File No.:	00018	Cert. No.: 00002
Sequential File No.:	00018	Lin. Date: 12/11/2008
		Lin. No.: 00002
		Soln. Date: 02/01/2009
		Soln. No.: 00007
		File Date: 02/01/2009
Calibrating Unit:	4	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXD S3-0192
Solution Control Lot:	08A049	Expires: 01/09/2010
		Bottle No.: 1129

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:56S		
Control 1 EC	0.104%	08:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.104%	08:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57S		
Control 2 EC	0.104%	08:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	08:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 3 EC	0.103%	08:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	08:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHNEZLER

First Name: EDWARD

MI: J

Signature: *TPRE J. Schnezler #5515*

Badge No.: 5515

Date: 02/01/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0012
Location:	NJSP LAKE HOPATCONG		
Calibration File No.:	00013	Calib. Date:	12/11/2008
Certification File No.:	00014	Cert. Date:	12/11/2008
Linearity File No.:	00015	Lin. Date:	12/11/2008
Solution File No.:	00019	Soln. Date:	03/05/2009
Sequential File No.:	00019	File Date:	03/05/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0179
Solution Control Lot:	08A049	Expires:	01/09/2010
		Bottle No.:	1148

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:51S		
Control 1 EC	0.102%	16:52S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:52S		
Control 2 EC	0.101%	16:53S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	16:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:54S		
Control 3 EC	0.101%	16:55S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	16:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDXAP2-128 (CAR)

Changed By:

Last Name: ROUSHINKO

First Name: CRAIG

MI: A.

Signature: TPR II C.R. #6046

Badge No.: 6046

Date: 03/05/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0012
Location:	NJSP LAKE HOPATCONG	
Calibration File No.:	00013	Calib. Date: 12/11/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 12/11/2008
Solution File No.:	00020	Cert. No.: 00002
Sequential File No.:	00020	Lin. Date: 12/11/2008
		Lin. No.: 00002
		Soln. Date: 04/02/2009
		Soln. No.: 00009
		File Date: 04/02/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0179
Solution Control Lot:	08A049	Expires: 01/09/2010
		Bottle No.: 1139

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:43S		
Control 1 EC	0.101%	13:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45S		
Control 2 EC	0.101%	13:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 3 EC	0.101%	13:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHNEZLER

First Name: EDWARD

MI: J

Signature: 

Badge No.: 5515

Date: 04/02/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0012
Location:	NJSP LAKE HOPATCONG	
Calibration File No.:	00021	Calib. Date: 05/26/2009
Certification File No.:	00022	Calib. No.: 00004
Linearity File No.:	00023	Cert. Date: 05/26/2009
Solution File No.:	00024	Cert. No.: 00003
Sequential File No.:	00024	Lin. Date: 05/26/2009
		Lin. No.: 00003
		Soln. Date: 05/26/2009
		Soln. No.: 00010
		File Date: 05/26/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0179
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0189

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:36D		
Control 1 EC	0.101%	13:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38D		
Control 2 EC	0.101%	13:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:39D		
Control 3 EC	0.101%	13:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDXAP2-128 (CAR)

Changed By:

Last Name: ROUSHINKO

First Name: CRAIG

MI: A.

Signature: TPR II C.R. #6046

Badge No.: 6046

Date: 05/26/2009

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0012
 Location: NJSP LAKE HOPATCONG
 Calibration File No.: 00021 Calib. Date: 05/26/2009 Calib. No.: 00004
 Certification File No.: 00022 Cert. Date: 05/26/2009 Cert. No.: 00003
 Linearity File No.: 00023 Lin. Date: 05/26/2009 Lin. No.: 00003
 Solution File No.: 00025 Soln. Date: 06/26/2009 Soln. No.: 00011
 Sequential File No.: 00025 File Date: 06/26/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0179
 Control Solution %: 0.100% Expires: 03/07/2010
 Solution Control Lot: 08C051 Bottle No.: 0178

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:47D		
Control 1 EC	0.100%	12:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:49D		
Control 2 EC	0.099%	12:50D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:50D		
Control 3 EC	0.099%	12:51D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GIARNIERI

First Name: NICHOLAS

MI: -

Signature:

Badge No.: 6081

Date: 06/26/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0012
Location:	NJSP LAKE HOPATCONG	
Calibration File No.:	00021	Calib. Date: 05/26/2009
Certification File No.:	00022	Calib. No.: 00004
Linearity File No.:	00023	Cert. Date: 05/26/2009
Solution File No.:	00027	Cert. No.: 00003
Sequential File No.:	00027	Lin. Date: 05/26/2009
		Lin. No.: 00003
		Soln. Date: 08/10/2009
		Soln. No.: 00013
		File Date: 08/10/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0179
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0177

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:16D		
Control 1 EC	-.---%	11:17D	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	11:17D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHNEZLER

First Name: EDWARD

MI: J

Signature: *PRE E. J. Schnezler 5575*

Badge No.: 5515

Date: 08/10/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0012
Location:	NJSP LAKE HOPATCONG	
Calibration File No.:	00021	Calib. Date: 05/26/2009
Certification File No.:	00022	Calib. No.: 00004
Linearity File No.:	00023	Cert. Date: 05/26/2009
Solution File No.:	00029	Cert. No.: 00003
Sequential File No.:	00029	Lin. Date: 05/26/2009
		Lin. No.: 00003
		Soln. Date: 11/22/2009
		Soln. No.: 00015
		File Date: 11/22/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0179
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0182

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:20S		
Control 1 EC	0.100%	08:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:22S		
Control 2 EC	0.100%	08:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23S		
Control 3 EC	0.100%	08:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHNEZLER

First Name: EDWARD

MI: J

Signature: *TRIE J. Schnezler #5515*

Badge No.: 5515

Date: 11/22/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0012
Location:	NJSP LAKE HOPATCONG	
Calibration File No.:	00030	Calib. Date: 04/12/2010
Certification File No.:	00031	Calib. No.: 00005
Linearity File No.:	00032	Cert. Date: 04/12/2010
Solution File No.:	00033	Cert. No.: 00004
Sequential File No.:	00033	Lin. Date: 04/12/2010
		Lin. No.: 00004
		Soln. Date: 04/12/2010
		Soln. No.: 00016
		File Date: 04/12/2010

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDSC S3-0003
Solution Control Lot:	09G067	Expires: 07/01/2011
		Bottle No.: 0743

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:45D		
Control 1 EC	0.104%	10:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	10:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		
Control 2 EC	0.103%	10:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.104%	10:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48D		
Control 3 EC	0.103%	10:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	10:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUNP2-451 (CAR)

Changed By:

Last Name: ROUSHINKO

First Name: CRAIG

MI: A.

Signature: TPR II [Signature] #6046

Badge No.: 6046

Date: 04/12/2010

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0012
Location:	NJSP LAKE HOPATCONG	
Calibration File No.:	00030	Calib. Date: 04/12/2010
Certification File No.:	00031	Calib. No.: 00005
Linearity File No.:	00032	Cert. Date: 04/12/2010
Solution File No.:	00035	Cert. No.: 00004
Sequential File No.:	00035	Lin. Date: 04/12/2010
		Lin. No.: 00004
		Soln. Date: 07/02/2010
		Soln. No.: 00018
		File Date: 07/02/2010

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDSC S3-0003
Control Solution %: 0.100%		Expires: 07/01/2011
Solution Control Lot: 09G067		Bottle No.: 0731

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:17D		
Control 1 EC	0.103%	13:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	13:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18D		
Control 2 EC	0.102%	13:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20D		
Control 3 EC	0.101%	13:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHNEZLER

First Name: EDWARD

MI: J

Signature: *TPRT E.J. Schnozler #5515*

Badge No.: 5515

Date: 07/02/2010