

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00013	Calib. Date: 11/12/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 11/12/2008
Solution File No.:	00016	Cert. No.: 00001
Sequential File No.:	00016	Lin. Date: 11/12/2008
		Lin. No.: 00001
		Soln. Date: 11/12/2008
		Soln. No.: 00002
		File Date: 11/12/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDSC S3-0003
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0904

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:01S		
Control 1 EC	0.101%	15:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:03S		
Control 2 EC	0.100%	15:03S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:04S		
Control 3 EC	0.099%	15:04S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*TEMPERATURE PROBE SERIAL NUMBER: DDWAP2-049 jso*

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_

*Jpc. JSD # 6027*

Badge No.: 6027

Date: 11/12/2008

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00013	Calib. Date: 11/12/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 11/12/2008
Solution File No.:	00017	Cert. No.: 00001
Sequential File No.:	00017	Lin. Date: 11/12/2008
		Lin. No.: 00001
		Soln. Date: 01/01/2009
		Soln. No.: 00003
		File Date: 01/01/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDSC S3-0003
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 1400

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:22S		
Control 1 EC	0.096%	17:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:23S		
Control 2 EC	0.096%	17:24S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:24S		
Control 3 EC	0.096%	17:25S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: RODEMANN

First Name: PAUL

MI: R

Signature: \_\_\_\_\_



Badge No.: 6011

Date: 01/01/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00013	Calib. Date: 11/12/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 11/12/2008
Solution File No.:	00018	Cert. No.: 00001
Sequential File No.:	00018	Lin. Date: 11/12/2008
		Lin. No.: 00001
		Soln. Date: 02/03/2009
		Soln. No.: 00004
		File Date: 02/03/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDSC S3-0003
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 9020

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:10S		
Control 1 EC	0.103%	17:11S	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.113%	17:11S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.009%	17:12S		
Control 2 EC	0.102%	17:13S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14S		
Control 3 EC	0.099%	17:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:15S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: **RODEMANN**

First Name: **PAUL**

MI: **R**

Signature:  6011

Badge No.: **6011**

Date: **02/03/2009**

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00013	Calib. Date: 11/12/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 11/12/2008
Solution File No.:	00019	Cert. No.: 00001
Sequential File No.:	00019	Lin. Date: 11/12/2008
		Lin. No.: 00001
		Soln. Date: 02/07/2009
		Soln. No.: 00005
		File Date: 02/07/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: 4
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 9030

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:50S		
Control 1 EC	-.---%	08:51S	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	08:51S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRUZZI

First Name: JOSEPH

MI: S

Signature: SSGT [Signature] 4563

Badge No.: 4563

Date: 02/07/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0016
Location:	NJSP - OCEAN STATION		
Calibration File No.:	00013	Calib. Date:	11/12/2008
Certification File No.:	00014	Cert. No.:	00003
Linearity File No.:	00015	Lin. Date:	11/12/2008
Solution File No.:	00020	Soln. No.:	00001
Sequential File No.:	00020	Soln. Date:	02/07/2009
		File Date:	02/07/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	4
Solution Control Lot:	08C051	Expires:	03/07/2010
		Bottle No.:	0905

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:21S		
Control 1 EC	-.---%	11:22S	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	11:22S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name:	PETRUZZI	First Name:	JOSEPH	MI:	S
Signature:	<u>SSGT [Signature] 4563</u>			Badge No.:	4563
				Date:	02/07/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00013	Calib. Date: 11/12/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 11/12/2008
Solution File No.:	00021	Cert. No.: 00001
Sequential File No.:	00021	Lin. Date: 11/12/2008
		Lin. No.: 00001
		Soln. Date: 02/07/2009
		Soln. No.: 00007
		File Date: 02/07/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: 4
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0905

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:35S		
Control 1 EC	0.101%	12:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:36S		
Control 2 EC	0.098%	12:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38S		
Control 3 EC	0.099%	12:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRUZZI	First Name: JOSEPH	MI: S
Signature: <u>SSGT J. AB 4563</u>	Badge No.: 4563	Date: 02/07/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00013	Calib. Date: 11/12/2008
Certification File No.:	00014	Calib. No.: 00003
Linearity File No.:	00015	Cert. Date: 11/12/2008
Solution File No.:	00022	Cert. No.: 00001
Sequential File No.:	00022	Lin. Date: 11/12/2008
		Lin. No.: 00001
		Soln. Date: 03/12/2009
		Soln. No.: 00008
		File Date: 03/12/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDRF S3-0010
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0908

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:11S		
Control 1 EC	0.101%	16:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:13S		
Control 2 EC	0.101%	16:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:14S		
Control 3 EC	0.100%	16:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*TEMPERATURE PROBE SERIAL NUMBER: DDXAP2-032 jfb*

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: *Tpr. II [Signature] #6027*

Badge No.: 6027

Date: 03/12/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00023	Calib. Date: 04/06/2009
Certification File No.:	00024	Calib. No.: 00004
Linearity File No.:	00025	Cert. Date: 04/06/2009
Solution File No.:	00026	Cert. No.: 00002
Sequential File No.:	00026	Lin. Date: 04/06/2009
		Lin. No.: 00002
		Soln. Date: 04/06/2009
		Soln. No.: 00009
		File Date: 04/06/2009

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDRF S3-0010
Control Solution %: 0.100%		Expires: 03/07/2010
Solution Control Lot: 08C051		Bottle No.: 0913

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:19D		
Control 1 EC	0.101%	14:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:21D		
Control 2 EC	0.100%	14:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:22D		
Control 3 EC	0.100%	14:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDXAP2 - 032<sub>g</sub>

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: Tpr. II J. Dellanoce # 6027

Badge No.: 6027

Date: 04/06/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00023	Calib. Date: 04/06/2009
Certification File No.:	00024	Calib. No.: 00004
Linearity File No.:	00025	Cert. Date: 04/06/2009
Solution File No.:	00027	Cert. No.: 00002
Sequential File No.:	00027	Lin. Date: 04/06/2009
		Lin. No.: 00002
		Soln. Date: 06/30/2009
		Soln. No.: 00010
		File Date: 06/30/2009

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDRF S3-0010
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0909

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:00D		
Control 1 EC	0.103%	18:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:02D		
Control 2 EC	0.101%	18:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:03D		
Control 3 EC	0.100%	18:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: READER

First Name: SEAN

MI: K

Signature:  #6151

Badge No.: 6151

Date: 06/30/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00023	Calib. Date: 04/06/2009
Certification File No.:	00024	Calib. No.: 00004
Linearity File No.:	00025	Cert. Date: 04/06/2009
Solution File No.:	00028	Cert. No.: 00002
Sequential File No.:	00028	Lin. Date: 04/06/2009
		Lin. No.: 00002
		Soln. Date: 08/03/2009
		Soln. No.: 00011
		File Date: 08/03/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDRF S3-0010
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0906

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:59D		
Control 1 EC	-.---%	14:59D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	14:59D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRUZZI	First Name: JOSEPH	MI: S
Signature: <u>SSGT [Signature] 4563</u>		Badge No.: 4563
		Date: 08/03/2009



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00023	Calib. Date: 04/06/2009
Certification File No.:	00024	Calib. No.: 00004
Linearity File No.:	00025	Cert. Date: 04/06/2009
Solution File No.:	00030	Cert. No.: 00002
Sequential File No.:	00030	Lin. Date: 04/06/2009
		Lin. No.: 00002
		Soln. Date: 08/29/2009
		Soln. No.: 00013
		File Date: 08/29/2009


Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDRF S3-0010
Control Solution %: 0.100%		Expires: 03/07/2010
Solution Control Lot: 08C051		Bottle No.: 0907

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:16D		
Control 1 EC	0.101%	12:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17D		
Control 2 EC	0.099%	12:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18D		
Control 3 EC	0.100%	12:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: READER	First Name: TPR II	MI: S
Signature: <u> #6151</u>	Badge No.: 6151	Date: 08/29/2009

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0016  
 Location: NJSP - OCEAN STATION  
 Calibration File No.: 00031 Calib. Date: 09/29/2009 Calib. No.: 00005  
 Certification File No.: 00032 Cert. Date: 09/29/2009 Cert. No.: 00003  
 Linearity File No.: 00033 Lin. Date: 09/29/2009 Lin. No.: 00003  
 Solution File No.: 00034 Soln. Date: 09/29/2009 Soln. No.: 00014  
 Sequential File No.: 00034 File Date: 09/29/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDRF S3-0010  
 Control Solution %: 0.100% Expires: 03/07/2010  
 Solution Control Lot: 08C051 Bottle No.: 0914

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:29D		
Control 1 EC	0.100%	15:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30D		
Control 2 EC	0.098%	15:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32D		
Control 3 EC	0.099%	15:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*TEMPERATURE PROBE SERIAL NUMBER: DDXAP2 - 032*

**Changed By:**

Last Name: DELLANOCE First Name: JOSEPH MI: S

Signature: *Tpr. II J. Dellanoce # 6027* Badge No.: 6027  
 Date: 09/29/2009

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00031	Calib. Date: 09/29/2009
Certification File No.:	00032	Calib. No.: 00005
Linearity File No.:	00033	Cert. Date: 09/29/2009
Solution File No.:	00035	Cert. No.: 00003
Sequential File No.:	00035	Lin. Date: 09/29/2009
		Lin. No.: 00003
		Soln. Date: 11/24/2009
		Soln. No.: 00015
		File Date: 11/24/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDRF S3-0010
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0918

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:07S		
Control 1 EC	0.099%	08:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:08S		
Control 2 EC	0.098%	08:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:09S		
Control 3 EC	0.098%	08:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PETRUZZI

First Name: JOSEPH

MI: S

Signature: \_\_\_\_\_

*SSGT J. Petruzzi #4563*

Badge No.: 4563

Date: 11/24/2009

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0016  
 Location: NJSP - OCEAN STATION  
 Calibration File No.: 00031 Calib. Date: 09/29/2009 Calib. No.: 00005  
 Certification File No.: 00032 Cert. Date: 09/29/2009 Cert. No.: 00003  
 Linearity File No.: 00033 Lin. Date: 09/29/2009 Lin. No.: 00003  
 Solution File No.: 00036 Soln. Date: 01/14/2010 Soln. No.: 00016  
 Sequential File No.: 00036 File Date: 01/14/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDRF S3-0010  
 Control Solution %: 0.100% Expires: 03/07/2010  
 Solution Control Lot: 08C051 Bottle No.: 0919

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:28S		
Control 1 EC	-.---%	08:29S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	08:29S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PFEIFER

First Name: PETER

MI: J

Signature:  5766

Badge No.: 5766

Date: 01/14/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0016
Location:	NJSP - OCEAN STATION	
Calibration File No.:	00031	Calib. Date: 09/29/2009
Certification File No.:	00032	Calib. No.: 00005
Linearity File No.:	00033	Cert. Date: 09/29/2009
Solution File No.:	00037	Cert. No.: 00003
Sequential File No.:	00037	Lin. Date: 09/29/2009
		Lin. No.: 00003
		Soln. Date: 01/14/2010
		Soln. No.: 00017
		File Date: 01/14/2010

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDRF S3-0010
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0919

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:45S		
Control 1 EC	0.100%	12:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47S		
Control 2 EC	0.099%	12:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48S		
Control 3 EC	0.097%	12:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:50S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PFEIFER

First Name: PETER

MI: J

Signature:  5766

Badge No.: 5766

Date: 01/14/2010



# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0016  
 Location: NJSP - OCEAN STATION  
 Calibration File No.: 00031 Calib. Date: 09/29/2009 Calib. No.: 00005  
 Certification File No.: 00032 Cert. Date: 09/29/2009 Cert. No.: 00003  
 Linearity File No.: 00033 Lin. Date: 09/29/2009 Lin. No.: 00003  
 Solution File No.: 00039 Soln. Date: 03/16/2010 Soln. No.: 00019  
 Sequential File No.: 00039 File Date: 03/16/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAE-0021  
 Control Solution %: 0.100% Expires: 07/01/2011  
 Solution Control Lot: 09G067 Bottle No.: 0819

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:20S		
Control 1 EC	0.101%	12:20S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21S		
Control 2 EC	0.098%	12:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22S		
Control 3 EC	0.098%	12:23S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETRUZZI

First Name: JOSEPH

MI: S

Signature: SCOTT J. PETRUZZI #4563

Badge No.: 4563

Date: 03/16/2010



