



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0058
Location:	NJSP - NORTH WILDWOOD	Calib. No.:	00002
Calibration File No.:	00006	Cert. Date:	10/16/2008
Certification File No.:	00007	Lin. Date:	10/16/2008
Linearity File No.:	00008	Soln. Date:	11/20/2008
Solution File No.:	00010	File Date:	11/20/2008
Sequential File No.:	00010		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	4
Solution Control Lot:	08C051	Expires:	03/07/2010
		Bottle No.:	0813

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:48S		
Control 1 EC	0.097%	11:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:49S		
Control 2 EC	0.095%	11:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51S		
Control 3 EC	0.096%	11:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MASLOWSKI

First Name: MICHAEL

MI: C

Signature:

*[Handwritten Signature: Tpr. I M Maslowski #5478]*

Badge No.: 5478

Date: 11/20/2008

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0058
Location:	NJSP - NORTH WILDWOOD	Calib. No.:	00002
Calibration File No.:	00006	Cert. Date:	10/16/2008
Certification File No.:	00007	Lin. Date:	10/16/2008
Linearity File No.:	00008	Soln. Date:	11/20/2008
Solution File No.:	00011	File Date:	11/20/2008
Sequential File No.:	00011		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXD S3-0180
Solution Control Lot:	08C051	Expires:	03/07/2010
		Bottle No.:	0814

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:28S		
Control 1 EC	0.099%	13:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29S		
Control 2 EC	0.097%	13:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31S		
Control 3 EC	0.097%	13:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MASLOWSKI

First Name: MICHAEL

MI: C

Signature: Tpr. E M Maslowski #5478

Badge No.: 5478

Date: 11/20/2008

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0058
Location:	NJSP - NORTH WILDWOOD	
Calibration File No.:	00006	Calib. Date: 10/16/2008
Certification File No.:	00007	Calib. No.: 00002
Linearity File No.:	00008	Cert. Date: 10/16/2008
Solution File No.:	00012	Cert. No.: 00001
Sequential File No.:	00012	Lin. Date: 10/16/2008
		Lin. No.: 00001
		Soln. Date: 12/29/2008
		Soln. No.: 00005
		File Date: 12/29/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXD S3-0180
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0801

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:21S		
Control 1 EC	0.100%	10:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22S		
Control 2 EC	0.100%	10:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:24S		
Control 3 EC	0.099%	10:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**TEMPERATURE PROBE SERIAL NUMBER: DDUNP2 - 273ans**

**Changed By:**

Last Name: STANKS

First Name: ADAM

MI: N

Signature: TPP. I Adam N. Stanks #5403

Badge No.: 5403

Date: 12/29/2008

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0058
Location:	NJSP - NORTH WILDWOOD	
Calibration File No.:	00006	Calib. Date: 10/16/2008
Certification File No.:	00007	Calib. No.: 00002
Linearity File No.:	00008	Cert. Date: 10/16/2008
Solution File No.:	00013	Cert. No.: 00001
Sequential File No.:	00013	Lin. Date: 10/16/2008
		Lin. No.: 00001
		Soln. Date: 02/10/2009
		Soln. No.: 00006
		File Date: 02/10/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXD S3-0180
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0806

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:44S		
Control 1 EC	0.103%	09:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45S		
Control 2 EC	0.100%	09:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47S		
Control 3 EC	0.100%	09:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUNP2-273 *ams*

**Changed By:**

Last Name: STANKS	First Name: ADAM	MI: N
Signature: <u><i>TPR. I Adam N. Stanks #5403</i></u>	Badge No.: 5403	Date: 02/10/2009



# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0058  
 Location: NJSP - NORTH WILDWOOD  
 Calibration File No.: 00015 Calib. Date: 04/02/2009 Calib. No.: 00003  
 Certification File No.: 00016 Cert. Date: 04/02/2009 Cert. No.: 00002  
 Linearity File No.: 00017 Lin. Date: 04/02/2009 Lin. No.: 00002  
 Solution File No.: 00018 Soln. Date: 04/02/2009 Soln. No.: 00008  
 Sequential File No.: 00018 File Date: 04/02/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWF S3-0220  
 Control Solution %: 0.100% Expires: 03/07/2010  
 Solution Control Lot: 08C051 Bottle No.: 0817

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:31S	34.0°C	*** TEST PASSED ***
Control 1 EC	0.101%	10:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32S	34.0°C	*** TEST PASSED ***
Control 2 EC	0.099%	10:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34S	34.0°C	*** TEST PASSED ***
Control 3 EC	0.098%	10:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35S	34.0°C	*** TEST PASSED ***

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDx2P2-016

**Changed By:**

Last Name: STANKS First Name: ADAM MI: N

Signature: TPR. I Adam N. Stank #5403 Badge No.: 5403 Date: 04/02/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0058
Location:	NJSP - NORTH WILDWOOD	Calib. No.:	00003
Calibration File No.:	00015	Cert. Date:	04/02/2009
Certification File No.:	00016	Lin. Date:	04/02/2009
Linearity File No.:	00017	Soln. Date:	06/02/2009
Solution File No.:	00019	File Date:	06/02/2009
Sequential File No.:	00019		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWF S3-0220
Solution Control Lot:	08C051	Expires:	03/07/2010
		Bottle No.:	0807

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:41D		
Control 1 EC	0.100%	19:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:43D		
Control 2 EC	0.099%	19:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:44D		
Control 3 EC	0.099%	19:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MASLOWSKI

First Name: MICHAEL

MI: C

Signature:



Badge No.: 5478

Date: 06/02/2009



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0058
Location:	NJSP - NORTH WILDWOOD	
Calibration File No.:	00015	Calib. Date: 04/02/2009
Certification File No.:	00016	Calib. No.: 00003
Linearity File No.:	00017	Cert. Date: 04/02/2009
Solution File No.:	00023	Cert. No.: 00002
Sequential File No.:	00023	Lin. Date: 04/02/2009
		Lin. No.: 00002
		Soln. Date: 07/09/2009
		Soln. No.: 00011
		File Date: 07/09/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAE-0020
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0802

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:49D		
Control 1 EC	0.099%	10:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:51D		
Control 2 EC	0.096%	10:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52D		
Control 3 EC	0.097%	10:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MASLOWSKI

First Name: MICHAEL

MI: C

Signature: \_\_\_\_\_



Badge No.: 5478

Date: 07/09/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0058
Location:	NJSP - NORTH WILDWOOD	
Calibration File No.:	00015	Calib. Date: 04/02/2009
Certification File No.:	00016	Calib. No.: 00003
Linearity File No.:	00017	Cert. Date: 04/02/2009
Solution File No.:	00029	Cert. No.: 00002
Sequential File No.:	00029	Lin. Date: 04/02/2009
		Lin. No.: 00002
		Soln. Date: 07/28/2009
		Soln. No.: 00012
		File Date: 07/28/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAE-0020
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0818

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:48D		
Control 1 EC	0.099%	14:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:49D		
Control 2 EC	0.097%	14:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:50D		
Control 3 EC	0.097%	14:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MASLOWSKI

First Name: MICHAEL

MI: C

Signature: \_\_\_\_\_



Badge No.: 5478

Date: 07/28/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0058
Location:	NJSP - NORTH WILDWOOD	Calib. No.:	00004
Calibration File No.:	00035	Cert. Date:	09/29/2009
Certification File No.:	00036	Lin. Date:	09/29/2009
Linearity File No.:	00037	Soln. Date:	09/29/2009
Solution File No.:	00038	File Date:	09/29/2009
Sequential File No.:	00038		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAE-0020
Solution Control Lot:	08C051	Expires:	03/07/2010
		Bottle No.:	0805

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:19D		
Control 1 EC	0.099%	11:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:21D		
Control 2 EC	0.098%	11:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D		
Control 3 EC	0.098%	11:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*Temperature Probe Serial Number: DD XK Pa-33425*

**Changed By:**

Last Name: SULLIVAN

First Name: FRANCIS

MI: -

Signature: *Francis J. Sullivan #5103*

Badge No.: 5103

Date: 09/29/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0058
Location:	NJSP - NORTH WILDWOOD	
Calibration File No.:	00035	Calib. Date: 09/29/2009
Certification File No.:	00036	Calib. No.: 00004
Linearity File No.:	00037	Cert. Date: 09/29/2009
Solution File No.:	00040	Cert. No.: 00003
Sequential File No.:	00040	Lin. Date: 09/29/2009
		Lin. No.: 00003
		Soln. Date: 02/03/2010
		Soln. No.: 00016
		File Date: 02/03/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWF S3-0262
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0810

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:56S		
Control 1 EC	0.100%	12:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:57S		
Control 2 EC	0.097%	12:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59S		
Control 3 EC	0.099%	12:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*Temperature Probe Serial Number: DD 21K P2-44723*

**Changed By:**

Last Name: SULLIVAN	First Name: FRANCIS	MI: -
Signature: <u><i>TP I J Sullivan #5103</i></u>	Badge No.: 5103	Date: 02/03/2010

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWC-0058  
**Location:** NJSP - NORTH WILDWOOD  
**Calibration File No.:** 00041 **Calib. Date:** 03/18/2010 **Calib. No.:** 00005  
**Certification File No.:** 00042 **Cert. Date:** 03/18/2010 **Cert. No.:** 00004  
**Linearity File No.:** 00043 **Lin. Date:** 03/18/2010 **Lin. No.:** 00004  
**Solution File No.:** 00044 **Soln. Date:** 03/18/2010 **Soln. No.:** 00017  
**Sequential File No.:** 00044 **File Date:** 03/18/2010

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWF S3-0262  
**Control Solution %:** 0.100% **Expires:** 05/22/2011  
**Solution Control Lot:** 09E066 **Bottle No.:** 0875

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:11S	34.0°C	
Control 1 EC	0.100%	15:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:12S		
Control 2 EC	0.099%	15:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14S		
Control 3 EC	0.098%	15:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*Temperature Probe Serial Number: DDXK12-44725*

**Changed By:**

Last Name: SULLIVAN

First Name: FRANCIS

MI: -

Signature: *Tpr T.J. Sullivan #5103*

Badge No.: 5103

Date: 03/18/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0058
Location:	NJSP - NORTH WILDWOOD	
Calibration File No.:	00041	Calib. Date: 03/18/2010
Certification File No.:	00042	Calib. No.: 00005
Linearity File No.:	00043	Cert. Date: 03/18/2010
Solution File No.:	00047	Cert. No.: 00004
Sequential File No.:	00047	Lin. Date: 03/18/2010
		Lin. No.: 00004
		Soln. Date: 05/26/2010
		Soln. No.: 00020
		File Date: 05/26/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWF S3-0262
Solution Control Lot:	09G067	Expires: 07/01/2011
		Bottle No.: 0328

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:50D		
Control 1 EC	0.097%	14:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	14:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		
Control 2 EC	0.096%	14:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	14:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:53D		
Control 3 EC	0.096%	14:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	14:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MASLOWSKI

First Name: MICHAEL

MI: C

Signature: \_\_\_\_\_



Badge No.: 5478

Date: 05/26/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0058
Location:	NJSP - NORTH WILDWOOD	
Calibration File No.:	00041	Calib. Date: 03/18/2010
Certification File No.:	00042	Calib. No.: 00005
Linearity File No.:	00043	Cert. Date: 03/18/2010
Solution File No.:	00048	Cert. No.: 00004
Sequential File No.:	00048	Lin. Date: 03/18/2010
		Lin. No.: 00004
		Soln. Date: 07/18/2010
		Soln. No.: 00021
		File Date: 07/18/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWF S3-0262
Solution Control Lot:	09G067	Expires: 07/01/2011
		Bottle No.: 0343

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:03D		
Control 1 EC	0.100%	07:03D	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	07:03D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	07:04D		
Control 2 EC	-.---%	---:---	N/A	
Control 2 IR	-.---%	---:---	N/A	
Ambient Air Blank	-.---%	---:---		
Control 3 EC	-.---%	---:---	N/A	
Control 3 IR	-.---%	---:---	N/A	
Ambient Air Blank	-.---%	---:---		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MULHOLLAND

First Name: SEAN

MI: G

Signature: \_\_\_\_\_



5383

Badge No.: 5383

Date: 07/18/2010

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0058
Location:	NJSP - NORTH WILDWOOD	
Calibration File No.:	00041	Calib. Date: 03/18/2010
Certification File No.:	00042	Calib. No.: 00005
Linearity File No.:	00043	Cert. Date: 03/18/2010
Solution File No.:	00049	Cert. No.: 00004
Sequential File No.:	00049	Lin. Date: 03/18/2010
		Lin. No.: 00004
		Soln. Date: 07/21/2010
		Soln. No.: 00022
		File Date: 07/21/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWF S3-0262
Solution Control Lot:	09G067	Expires: 07/01/2011
		Bottle No.: 0344

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:38D		
Control 1 EC	0.100%	17:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:39D		
Control 2 EC	0.097%	17:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:40D		
Control 3 EC	0.097%	17:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MULHOLLAND

First Name: SEAN

MI: G

Signature: \_\_\_\_\_



5383

Badge No.: 5383

Date: 07/21/2010