

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0064
Location:	NJSP - MONMOUTH STATION	
Calibration File No.:	00009	Calib. Date: 11/06/2008
Certification File No.:	00010	Cert. No.: 00001
Linearity File No.:	00011	Lin. Date: 11/06/2008
Solution File No.:	00012	Soln. No.: 00002
Sequential File No.:	00012	File Date: 11/06/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWF S3-0264
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 1271

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:40S		
Control 1 EC	0.100%	14:41S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:41S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41S		
Control 2 EC	0.098%	14:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:42S		
Control 3 EC	0.098%	14:43S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:43S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**TEMPERATURE PROBE SERIAL NUMBER: DDWAP2-105ams**

**Changed By:**

Last Name: STANKS

First Name: ADAM

MI: N

Signature: TPR. I Adam M. Stanks #5403

Badge No.: 5403

Date: 11/06/2008

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0064
Location:	NJSP - MONMOUTH STATION	
Calibration File No.:	00009	Calib. Date: 11/06/2008
Certification File No.:	00010	Calib. No.: 00002
Linearity File No.:	00011	Cert. Date: 11/06/2008
Solution File No.:	00013	Cert. No.: 00001
Sequential File No.:	00013	Lin. Date: 11/06/2008
		Lin. No.: 00001
		Soln. Date: 01/31/2009
		Soln. No.: 00003
		File Date: 01/31/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWF S3-0264
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 1272

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:13S		
Control 1 EC	0.098%	14:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15S		
Control 2 EC	0.098%	14:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:16S		
Control 3 EC	0.097%	14:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HYBBENETH

First Name: SEAN

MI: S

Signature: \_\_\_\_\_

*Tr S. Hybbeneth* 5998

Badge No.: 5998

Date: 01/31/2009



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0064
Location:	NJSP - MONMOUTH STATION	
Calibration File No.:	00009	Calib. Date: 11/06/2008
Certification File No.:	00010	Calib. No.: 00002
Linearity File No.:	00011	Cert. Date: 11/06/2008
Solution File No.:	00015	Cert. No.: 00001
Sequential File No.:	00015	Lin. Date: 11/06/2008
		Lin. No.: 00001
		Soln. No.: 00005
		File Date: 04/20/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXD S3-0192
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0934

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:45D		
Control 1 EC	0.099%	10:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		
Control 2 EC	0.097%	10:47D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:47D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48D		
Control 3 EC	0.097%	10:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHNYDERITE

First Name: RICHARD

MI: T

Signature: \_\_\_\_\_

Badge No.: 4745

Date: 04/20/2009





# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0064
<b>Location:</b>	NJSP - MONMOUTH STATION	
<b>Calibration File No.:</b>	00016	<b>Calib. Date:</b> 05/01/2009
<b>Certification File No.:</b>	00017	<b>Calib. No.:</b> 00003
<b>Linearity File No.:</b>	00018	<b>Cert. Date:</b> 05/01/2009
<b>Solution File No.:</b>	00025	<b>Cert. No.:</b> 00002
<b>Sequential File No.:</b>	00025	<b>Lin. Date:</b> 05/01/2009
		<b>Lin. No.:</b> 00002
		<b>Soln. Date:</b> 09/12/2009
		<b>Soln. No.:</b> 00011
		<b>File Date:</b> 09/12/2009

<b>Calibrating Unit:</b> WET	<b>Model No.:</b> CU-34	<b>Serial No.:</b> DDXD S3-0192
<b>Control Solution %:</b> 0.100%		<b>Expires:</b> 03/07/2010
<b>Solution Control Lot:</b> 08C051		<b>Bottle No.:</b> 0583

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:02D		
Control 1 EC	0.100%	21:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		
Control 2 EC	0.098%	21:04D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05D		
Control 3 EC	0.098%	21:05D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BROBST

First Name: KARL

MI: J

Signature: \_\_\_\_\_



Badge No.: 5429

Date: 09/12/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWC-0064
Location:	NJSP - MONMOUTH STATION	
Calibration File No.:	00026	Calib. Date: 10/14/2009
Certification File No.:	00027	Calib. No.: 00004
Linearity File No.:	00028	Cert. Date: 10/14/2009
Solution File No.:	00029	Cert. No.: 00003
Sequential File No.:	00029	Lin. Date: 10/14/2009
		Lin. No.: 00003
		Soln. Date: 10/14/2009
		Soln. No.: 00012
		File Date: 10/14/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXD S3-0192
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0589

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:49D		
Control 1 EC	0.101%	11:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51D		
Control 2 EC	0.099%	11:51D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52D		
Control 3 EC	0.100%	11:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: \_\_\_\_\_

Badge No.: 5792

Date: 10/14/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0064
Location:	NJSP - MONMOUTH STATION	Calib. No.:	00004
Calibration File No.:	00026	Cert. Date:	10/14/2009
Certification File No.:	00027	Lin. Date:	10/14/2009
Lincarity File No.:	00028	Soln. Date:	11/21/2009
Solution File No.:	00030	File Date:	11/21/2009
Sequential File No.:	00030		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXD S3-0192
Solution Control Lot:	08C051	Expires:	03/07/2010
		Bottle No.:	0578

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:31S		
Control 1 EC	0.101%	07:31S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:32S		
Control 2 EC	0.100%	07:33S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:33S		
Control 3 EC	0.100%	07:34S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:35S		

▀ All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SCHNYDERITE

First Name: SGT41

MI: T

Signature: \_\_\_\_\_



Badge No.: 4745

Date: 11/21/2009

## Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0064
<b>Location:</b>	NJSP - MONMOUTH STATION		
<b>Calibration File No.:</b>	00026	<b>Calib. Date:</b>	10/14/2009
<b>Certification File No.:</b>	00027	<b>Cert. No.:</b>	00004
<b>Linearity File No.:</b>	00028	<b>Cert. Date:</b>	10/14/2009
<b>Solution File No.:</b>	00031	<b>Lin. Date:</b>	10/14/2009
<b>Sequential File No.:</b>	00031	<b>Lin. No.:</b>	00003
		<b>Soln. Date:</b>	12/29/2009
		<b>Soln. No.:</b>	00014
		<b>File Date:</b>	12/29/2009
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDXD S3-0192
<b>Solution Control Lot:</b>	08C051	<b>Expires:</b>	03/07/2010
		<b>Bottle No.:</b>	0578

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:35S		
Control 1 EC	0.103%	10:35S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36S		
Control 2 EC	0.100%	10:37S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.118%	10:37S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.010%	10:37S		
Control 3 EC	0.100%	10:38S	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.091%	10:38S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:39S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DUFFY

First Name: JAMES

MI: 4

Signature: \_\_\_\_\_

Badge No.: 5439

Date: 12/29/2009

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0064
Location:	NJSP - MONMOUTH STATION		
Calibration File No.:	00026	Calib. Date:	10/14/2009
Certification File No.:	00027	Cert. Date:	10/14/2009
Linearity File No.:	00028	Lin. Date:	10/14/2009
Solution File No.:	00033	Soln. Date:	02/11/2010
Sequential File No.:	00033	File Date:	02/11/2010
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXD S3-0187
Solution Control Lot:	09A062	Expires:	01/15/2011
		Bottle No.:	0958

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:36S		
Control 1 EC	0.102%	19:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:37S		
Control 2 EC	0.100%	19:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:39S		
Control 3 EC	0.098%	19:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DAWS P2 - 053 

**Changed By:**  
Last Name: SNYDER First Name: THOMAS MI: J.

Signature: TR. I Thomas J. Snyder #5792 Badge No.: 5792  
Date: 02/11/2010

