

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWD-0102
Location:	N.J.S.P. SPORTS COMPLEX-	
Calibration File No.:	00007	Calib. Date: 03/04/2009
Certification File No.:	00008	Calib. No.: 00002
Linearity File No.:	00009	Cert. Date: 03/04/2009
Solution File No.:	00011	Cert. No.: 00001
Sequential File No.:	00011	Lin. Date: 03/04/2009
		Lin. No.: 00001
		Soln. Date: 03/27/2009
		Soln. No.: 00003
		File Date: 03/27/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWF S3-0258
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0223

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:20S		
Control 1 EC	0.102%	10:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22S		
Control 2 EC	0.099%	10:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23S		
Control 3 EC	0.100%	10:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ARROYO

First Name: MARCOS

MI: A.

Signature: _____

Badge No.: 6291

Date: 03/27/2009

Calibrating Unit New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWD-0102
Location:	N.J.S.P. SPORTS COMPLEX	
Calibration File No.:	00007	Calib. Date: 03/04/2009
Certification File No.:	00008	Calib. No.: 00002
Linearity File No.:	00009	Cert. Date: 03/04/2009
Solution File No.:	00012	Cert. No.: 00001
Sequential File No.:	00012	Lin. Date: 03/04/2009
		Lin. No.: 00001
		Soln. Date: 05/01/2009
		Soln. No.: 00004
		File Date: 05/01/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWF S3-0258
Solution Control Lot:	08C051	Expires: 03/07/2010
		Bottle No.: 0217

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:52D		
Control 1 EC	0.102%	10:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54D		
Control 2 EC	0.100%	10:54D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		
Control 3 EC	0.099%	10:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NARDONE	First Name: EMILIO	MI: G.
Signature: <u> #5157</u>	Badge No.: 5757	Date: 05/01/2009

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWD-0102
Location:	N.J.S.P. SPORTS COMPLEX	
Calibration File No.:	00027	Calib. Date: 05/05/2010
Certification File No.:	00028	Calib. No.: 00004
Linearity File No.:	00029	Cert. Date: 05/05/2010
Solution File No.:	00030	Cert. No.: 00003
Sequential File No.:	00030	Lin. Date: 05/05/2010
		Lin. No.: 00003
		Soln. Date: 05/05/2010
		Soln. No.: 00016
		File Date: 05/05/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWE S3-0204
Solution Control Lot:	09G067	Expires: 07/01/2011
		Bottle No.: 1398

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:29D		
Control 1 EC	0.102%	14:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30D		
Control 2 EC	0.100%	14:31D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31D		
Control 3 EC	0.101%	14:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL # DDWJP2-203 (RMD)


Changed By:

Last Name: DEMLER

First Name: ROBERT

MI: M.

Signature: _____



Badge No.: 6073

Date: 05/05/2010